




High Temperature/ pH Amitriptyline Stability using HALO® Elevate

369



TEST CONDITIONS:

Column: HALO 120 Å ELV C18, 2.7 μm , 2.1 x 50 mm
 Part Number: 92272-402
 Mobile Phase A: 95/5 10mM Ammonium Bicarbonate,
 pH:10/ Acetonitrile
 Mobile Phase B: Acetonitrile
 Gradient:

Time	%B
0.0	0
2.5	95
3.0	95
3.1	0
5.0	0

Flow Rate: 0.8 mL/min
 Back Pressure: 220 bar
 Temperature: 60 °C
 Detection: UV/PDA, 260 nm
 Injection Volume: 1.0 μL
 LC System: Shimadzu Nexera X2

PEAK IDENTITIES:

1. Uracil
2. Acenaphthene
3. Amitriptyline

A high temperature/ high pH stability run of a common tricyclic antidepressant (amitriptyline) is achieved using a HALO® Elevate C18 column. Less than a 1% change in retention is achieved over 20,000 column volumes.

INNOVATION YOU CAN TRUST,
 PERFORMANCE YOU CAN RELY ON



via neera 8/a 20141 Milano - tel. +39 02.8954201 fax +39 02.89542022
 www.cps.it - cps@cps.it

CPS
 Analitica
 for Chemistry